Topilene® R530

Polypropylene Homopolymer Hyosung Corporation

Message:

Characteristics:

Transparency

FDA

DMF 21499

USP Class VI

SFDA

Applications:

IV bottle

General Information			
Additive	High Slip		
Features	High Clarity		
	Homopolymer		
Uses	Medical Devices		
	Medical/Healthcare Applications		
Agency Ratings	DMF 21499		
	FDA Unspecified Rating		
	USP Class VI		
Physical	Nominal Value	Unit	Test Method
Melt Mass-Flow Rate (MFR) (230°C/2.16			
kg)	6.0	g/10 min	ASTM D1238
Mechanical	Nominal Value	Unit	Test Method
Flexural Modulus	932	MPa	ASTM D790

The information and data on this page are provided by manufacturers and document providers. SHANGHAI SUSHENG assumes no legal liability. It is strongly recommended to verify all technical data with material suppliers before final material selection. All rights belong to the original authors. If any infringement occurs, please contact us immediately.

Unit

J/m

Test Method

ASTM D256

Nominal Value

59

Recommended distributors for this material

Susheng Import & Export Trading Co.,Ltd.

Tel: +86 21 5895 8519

Impact

Unnotched Izod Impact (23°C)

Phone: +86 13424755533 Email: sales@su-jiao.com

No. 215, Lianhe North Road, Fengxian District, Shanghai, China

